
Midmark M11 - v1.0.0
Total Cycles: 1885

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	128.0 F	0.0
2:00	129.8 F	0.0
4:00	150.8 F	0.7
6:00	179.5 F	3.6
8:00	203.3 F	7.2
10:00	216.9 F	11.9
12:00	222.0 F	6.0
14:00	233.5 F	10.7
16:00	233.8 F	7.3
18:00	244.3 F	13.1
20:00	242.6 F	11.4
22:00	255.1 F	17.8
24:00	266.0 F	24.4

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.2
1:00	272.7 F	29.0
2:00	271.7 F	28.1
3:00	272.4 F	29.0
4:00	272.3 F	28.6

Min 271.4 F 28.0
Max 272.7 F 29.1

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	186.5 F	0.0
5:00	198.9 F	0.0
10:00	206.8 F	0.0
15:00	213.0 F	0.0
20:00	215.1 F	0.0
25:00	219.7 F	0.0
30:00	223.9 F	0.0

Drying Complete

FILLING: 0:49
HEATING: 25:45
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

04/28/2022 12:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____