
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001586
16:26 11/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:28
PRESSURIZING 7:30
STERILIZING 9:16
136.4 C 223kPa 9:16
136.5 C 223kPa 9:46
136.3 C 222kPa 10:16
136.2 C 221kPa 10:46
136.0 C 220kPa 11:16
136.4 C 218kPa 11:46
136.4 C 223kPa 12:16
136.3 C 222kPa 12:46
VENTING 12:46
AIR DRYING 13:53
CYCLE COMPLETE 15:28

Digital Signature #
5360029CCAB193C6
