
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000227
9:48 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:25
STERILIZING 5:51
136.4 C 232kPa 5:51
136.4 C 233kPa 6:21
136.3 C 233kPa 6:51
136.4 C 232kPa 7:21
136.3 C 233kPa 7:51
136.4 C 233kPa 8:21
136.4 C 233kPa 8:51
136.4 C 233kPa 9:21
136.4 C 233kPa 9:21
VENTING 9:21
AIR DRYING 10:20
CYCLE COMPLETE 20:02

Digital Signature #
A436F1A22F2DCDB8
