
Midmark M11 - v1.0.5
Total Cycles: 2161

Name: RITTER2
Sterilizer ID: V2313228
03/20/2023 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 169.0 F 3.2
4:00 177.7 F 7.0
6:00 200.3 F 12.1
8:00 219.0 F 6.8
10:00 231.3 F 12.3
12:00 235.4 F 9.8
14:00 236.4 F 8.8
16:00 249.4 F 15.5
18:00 261.6 F 22.6
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.6
2:00 271.7 F 29.5
3:00 272.3 F 29.5
4:00 271.3 F 29.3

Min 271.2 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 230.4 F 0.0
10:00 225.8 F 0.0
15:00 227.3 F 0.0
20:00 213.1 F 0.0
25:00 201.2 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

03/20/2023 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____