
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000765
7:01 09/11/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:28
136.0 C 220kPa 4:28
136.3 C 223kPa 4:58
136.2 C 223kPa 5:28
136.4 C 221kPa 5:58
136.4 C 224kPa 6:28
136.2 C 223kPa 6:58
136.3 C 223kPa 7:28
136.3 C 222kPa 7:58
136.3 C 222kPa 7:58
VENTING 7:58
AIR DRYING 8:53
CYCLE COMPLETE 9:02

Digital Signature #
9C4CFA9A7E921464
