
Midmark M11 - v1.0.0
Total Cycles: 2959

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2023 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 78.8 F 0.0
4:00 83.1 F 0.0
6:00 109.4 F 0.2
8:00 142.4 F 1.3
10:00 173.6 F 3.5
12:00 196.0 F 6.0
14:00 210.0 F 9.4
16:00 216.9 F 10.8
18:00 223.4 F 7.4
20:00 234.3 F 11.8
22:00 234.0 F 8.2
24:00 242.8 F 12.0
26:00 243.7 F 12.1
28:00 255.3 F 18.1
30:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.5
1:00 272.5 F 29.1
2:00 271.6 F 28.6
3:00 272.1 F 28.8
4:00 272.1 F 29.1

Min 271.5 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 202.9 F 0.0
10:00 200.5 F 0.0
15:00 196.6 F 0.0
20:00 194.4 F 0.0
25:00 198.8 F 0.0
30:00 223.5 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 31:37
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

03/01/2023 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____