

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007472  
8:29 03/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:09  
STERILIZING 4:17  
136.0 C 220kPa 4:17  
136.2 C 219kPa 4:47  
136.2 C 223kPa 5:17  
136.4 C 223kPa 5:47  
136.2 C 224kPa 6:17  
136.4 C 224kPa 6:47  
136.3 C 222kPa 7:17  
136.3 C 223kPa 7:47  
136.3 C 223kPa 7:47  
VENTING 7:47  
AIR DRYING 8:43  
CYCLE COMPLETE 9:05

-----  
Digital Signature #  
90D9BA28B1103E81  
-----