
Midmark M11 - v1.0.0
Total Cycles: 4807

Name: RITTER 1
Sterilizer ID: V2178198
08/12/2024 05:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 181.1 F 4.3
4:00 223.6 F 9.6
6:00 222.7 F 5.4
8:00 236.2 F 10.9
10:00 233.4 F 8.0
12:00 240.1 F 10.2
14:00 247.0 F 13.6
16:00 260.3 F 20.9
18:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.2 F 28.4
2:00 272.5 F 28.9
3:00 271.4 F 28.0
4:00 272.6 F 28.7

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 234.9 F 0.0
10:00 214.0 F 0.0
15:00 221.0 F 0.0
20:00 239.1 F 0.0
25:00 229.8 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:28
CYCLE COMPLETE

08/12/2024 06:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____