

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 004369  
11:20 15/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:29  
STERILIZING 3:19  
136.0 C 220kPa 3:19  
136.4 C 223kPa 3:49  
136.4 C 223kPa 4:19  
136.1 C 223kPa 4:49  
136.2 C 221kPa 5:19  
136.4 C 223kPa 5:49  
136.4 C 224kPa 6:19  
136.4 C 225kPa 6:49  
136.4 C 225kPa 6:49  
VENTING 6:49  
AIR DRYING 7:38  
CYCLE COMPLETE 7:46

-----  
Digital Signature #  
4B389DB60A78A664  
-----