

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3414

Name: RITTER2  
Sterilizer ID: V2313228  
02/27/2024 09:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 72.2 F 0.0  
4:00 97.9 F 0.0  
6:00 101.3 F 0.7  
8:00 144.0 F 3.4  
10:00 169.7 F 6.4  
12:00 191.5 F 9.8  
14:00 206.7 F 8.4  
16:00 217.4 F 6.7  
18:00 228.2 F 10.7  
20:00 230.8 F 7.8  
22:00 239.7 F 11.8  
24:00 237.4 F 9.8  
26:00 250.2 F 15.9  
28:00 260.8 F 22.0  
30:00 270.8 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.6  
2:00 271.5 F 29.2  
3:00 272.6 F 29.6  
4:00 271.5 F 29.0

Min 271.3 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 199.4 F 0.0  
10:00 199.0 F 0.0  
15:00 189.7 F 0.0  
20:00 186.5 F 0.0  
25:00 184.4 F 0.0  
30:00 182.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:47  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:08:52  
CYCLE COMPLETE

02/27/2024 10:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_