
Midmark M11 - v1.0.0
Total Cycles: 4472

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2024 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 175.5 F 3.3
4:00 189.8 F 6.9
6:00 206.7 F 11.6
8:00 219.6 F 5.5
10:00 233.9 F 9.9
12:00 237.7 F 9.0
14:00 241.7 F 12.1
16:00 240.3 F 10.7
18:00 252.9 F 16.9
20:00 263.8 F 23.3
22:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.2
2:00 271.9 F 28.6
3:00 271.2 F 28.5
4:00 272.3 F 29.2

Min 271.1 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 229.7 F 0.0
10:00 212.5 F 0.0
15:00 203.0 F 0.0
20:00 200.5 F 0.0
25:00 211.0 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

04/18/2024 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____