
Midmark M11 - v1.0.0
Total Cycles: 1010

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2021 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.4 F 0.0
2:00 183.7 F 3.2
4:00 202.5 F 7.2
6:00 219.0 F 12.4
8:00 230.2 F 7.6
10:00 238.9 F 13.1
12:00 239.9 F 10.6
14:00 237.1 F 9.0
16:00 251.7 F 16.0
18:00 263.7 F 22.9
20:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.6 F 28.8
2:00 271.2 F 28.0
3:00 272.5 F 29.0
4:00 272.0 F 28.4

Min 271.1 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 239.7 F 0.0
10:00 226.8 F 0.0
15:00 214.8 F 0.0
20:00 210.2 F 0.0
25:00 213.2 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

07/14/2021 11:44 AM
ACCEPTED: X
REJECTED:
Operator: U01