

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5035

Name: RITTER 1  
Sterilizer ID: V2178198  
10/29/2024 03:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 162.0 F 0.8  
4:00 183.3 F 4.6  
6:00 214.1 F 9.6  
8:00 224.0 F 6.8  
10:00 229.4 F 8.9  
12:00 241.1 F 10.7  
14:00 241.9 F 11.5  
16:00 239.3 F 9.8  
18:00 252.8 F 16.5  
20:00 264.5 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.2 F 28.6  
2:00 271.2 F 27.9  
3:00 272.3 F 28.7  
4:00 271.7 F 28.1  
  
Min 271.1 F 27.9  
Max 272.4 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.6 F 0.0  
5:00 223.1 F 0.0  
10:00 192.9 F 0.0  
15:00 243.6 F 0.0  
20:00 224.6 F 0.0  
25:00 235.5 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:00:31  
CYCLE COMPLETE

10/29/2024 04:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_