
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006584
8:22 09/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:12
STERILIZING 4:14
136.4 C 221kPa 4:14
136.1 C 221kPa 4:44
135.8 C 223kPa 5:14
136.4 C 222kPa 5:44
136.1 C 220kPa 6:14
136.4 C 223kPa 6:44
136.3 C 222kPa 7:14
136.0 C 222kPa 7:44
136.0 C 222kPa 7:44
VENTING 7:44
AIR DRYING 8:36
CYCLE COMPLETE 38:36

Digital Signature #
96936A2BA6995265
