
Midmark M11 - v1.0.0
Total Cycles: 4260

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2024 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.1 F 0.0
2:00 133.3 F 2.0
4:00 152.4 F 4.9
6:00 188.2 F 9.1
8:00 214.0 F 7.5
10:00 227.3 F 7.2
12:00 234.1 F 11.8
14:00 232.9 F 7.8
16:00 245.0 F 13.3
18:00 242.0 F 11.4
20:00 254.4 F 17.7
22:00 265.5 F 24.3
24:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 28.7
1:00 272.9 F 29.1
2:00 272.7 F 29.4
3:00 271.6 F 28.8
4:00 272.6 F 29.0

Min 271.6 F 28.6
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 236.4 F 0.0
10:00 220.4 F 0.0
15:00 211.0 F 0.0
20:00 210.4 F 0.0
25:00 207.1 F 0.0
30:00 206.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

02/01/2024 04:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____