
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006484
10:30 17/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:26
STERILIZING 3:21
136.4 C 222kPa 3:21
136.0 C 218kPa 3:51
136.4 C 224kPa 4:21
136.4 C 224kPa 4:51
136.2 C 221kPa 5:21
136.4 C 225kPa 5:51
136.4 C 221kPa 6:21
136.2 C 219kPa 6:51
136.2 C 219kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 10:23

Digital Signature #
314C66D3296985B2
