
Midmark M11 - v1.0.0
Total Cycles: 4222

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 157.2 F 1.7
4:00 166.5 F 4.8
6:00 189.1 F 8.8
8:00 213.5 F 9.2
10:00 227.4 F 7.0
12:00 233.8 F 11.7
14:00 233.1 F 7.9
16:00 245.6 F 13.3
18:00 242.8 F 11.4
20:00 254.7 F 17.4
22:00 265.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.6
2:00 271.4 F 27.9
3:00 272.6 F 28.7
4:00 271.8 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 228.6 F 0.0
10:00 190.5 F 0.0
15:00 185.2 F 0.0
20:00 182.4 F 0.0
25:00 179.1 F 0.0
30:00 178.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

01/24/2024 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____