

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000797  
7:07 04/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:05  
STERILIZING 3:03  
136.4 C 230kPa 3:03  
136.2 C 231kPa 3:33  
136.4 C 232kPa 4:03  
136.3 C 231kPa 4:33  
136.4 C 230kPa 5:03  
136.4 C 230kPa 5:33  
136.3 C 232kPa 6:03  
136.4 C 233kPa 6:33  
136.4 C 233kPa 6:33  
VENTING 6:33  
AIR DRYING 7:20  
CYCLE COMPLETE 7:37  
-----

Digital Signature #  
3F8F03B09BE9D74F  
-----