
Midmark M11 - v1.0.0
Total Cycles: 4584

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2024 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 180.9 F 2.3
4:00 181.2 F 6.1
6:00 194.2 F 10.2
8:00 223.1 F 6.7
10:00 221.3 F 8.7
12:00 233.3 F 13.4
14:00 236.1 F 10.1
16:00 237.0 F 8.8
18:00 247.3 F 14.2
20:00 258.6 F 20.4
22:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.7 F 29.4
2:00 271.7 F 28.8
3:00 271.4 F 28.9
4:00 272.6 F 29.4

Min 271.1 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 211.0 F 0.0
10:00 217.7 F 0.0
15:00 229.1 F 0.0
20:00 230.3 F 0.0
25:00 231.4 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

05/27/2024 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____