
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002464
12:37 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:12
STERILIZING 4:20
136.3 C 225kPa 4:20
136.4 C 226kPa 4:50
136.4 C 224kPa 5:20
136.4 C 225kPa 5:50
136.3 C 226kPa 6:20
136.4 C 226kPa 6:50
136.4 C 226kPa 7:20
136.2 C 225kPa 7:50
136.2 C 225kPa 7:50
VENTING 7:50
AIR DRYING 8:46
CYCLE COMPLETE 29:13

Digital Signature #
06D4E9CFDD069261
