

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3478

Name: RITTER 1  
Sterilizer ID: V2178198  
07/04/2023 10:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 177.4 F 2.5  
4:00 184.1 F 6.1  
6:00 206.9 F 11.1  
8:00 220.0 F 5.7  
10:00 231.3 F 11.0  
12:00 231.2 F 7.5  
14:00 244.4 F 13.5  
16:00 243.2 F 11.9  
18:00 255.6 F 18.3  
20:00 266.4 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.3 F 29.2  
2:00 272.4 F 28.9  
3:00 271.7 F 28.4  
4:00 271.3 F 28.4

Min 271.2 F 28.2  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 216.6 F 0.0  
10:00 237.5 F 0.0  
15:00 238.5 F 0.0  
20:00 230.8 F 0.0  
25:00 225.5 F 0.0  
30:00 225.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

07/04/2023 11:10 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_