
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000640
10:35 09/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:53
STERILIZING 3:58
136.0 C 227kPa 3:58
136.3 C 231kPa 4:28
136.3 C 232kPa 4:58
136.2 C 230kPa 5:28
136.2 C 232kPa 5:58
136.2 C 231kPa 6:28
135.9 C 229kPa 6:58
136.2 C 231kPa 7:28
136.2 C 231kPa 7:28
VENTING 7:28
AIR DRYING 8:18
CYCLE COMPLETE 23:18

Digital Signature #
8936F63AE32B483B
