
Midmark M11 - v1.0.5
Total Cycles: 701

Name: RITTER2
Sterilizer ID: V2313228
01/17/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.9 F 0.0
2:00 121.7 F 1.4
4:00 143.6 F 4.0
6:00 171.8 F 7.6
8:00 192.3 F 12.0
10:00 209.4 F 5.7
12:00 223.3 F 10.5
14:00 229.8 F 7.8
16:00 240.2 F 12.9
18:00 240.1 F 11.6
20:00 252.6 F 17.9
22:00 264.0 F 24.6
24:00 271.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.7
1:00 272.5 F 30.4
2:00 271.5 F 29.6
3:00 272.1 F 30.4
4:00 272.3 F 30.2

Min 271.2 F 29.6
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 212.1 F 0.0
10:00 225.9 F 0.0
15:00 214.6 F 0.0
20:00 222.0 F 0.0
25:00 231.0 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 24:07
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

01/17/2022 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____