
Midmark M11 - v1.0.5
Total Cycles: 3313

Name: RITTER2
Sterilizer ID: V2313228
01/24/2024 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 80.0 F 0.0
4:00 106.6 F 0.0
6:00 106.4 F 0.3
8:00 143.2 F 2.7
10:00 168.5 F 6.2
12:00 197.7 F 9.8
14:00 218.5 F 10.8
16:00 223.2 F 6.2
18:00 230.8 F 9.9
20:00 236.5 F 9.9
22:00 238.4 F 10.4
24:00 236.4 F 8.5
26:00 247.1 F 13.8
28:00 258.0 F 19.8
30:00 268.1 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.0
2:00 271.8 F 29.0
3:00 272.0 F 28.7
4:00 272.4 F 29.3

Min 271.4 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 222.9 F 0.0
10:00 220.1 F 0.0
15:00 207.5 F 0.0
20:00 204.7 F 0.0
25:00 209.9 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

01/24/2024 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____