

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000619  
14:59 30/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:16  
STERILIZING 4:24  
136.0 C 227kPa 4:24  
136.3 C 232kPa 4:54  
136.4 C 233kPa 5:24  
136.4 C 232kPa 5:54  
136.4 C 233kPa 6:24  
136.3 C 231kPa 6:54  
136.2 C 230kPa 7:24  
136.4 C 233kPa 7:54  
136.4 C 233kPa 7:54  
VENTING 7:54  
AIR DRYING 8:47  
CYCLE COMPLETE 23:47  
-----

Digital Signature #  
3D45BDD77C68FEB1  
-----