

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3518

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2023 11:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.5 F 0.0  
2:00 172.4 F 2.4  
4:00 183.1 F 5.5  
6:00 204.2 F 9.9  
8:00 225.5 F 7.9  
10:00 229.9 F 9.8  
12:00 234.9 F 8.8  
14:00 242.4 F 12.0  
16:00 240.1 F 10.3  
18:00 252.0 F 16.6  
20:00 263.2 F 22.9  
22:00 272.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 28.7  
2:00 272.0 F 29.1  
3:00 271.4 F 28.5  
4:00 272.6 F 29.6

Min 271.0 F 28.4  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 196.0 F 0.0  
10:00 195.4 F 0.0  
15:00 193.9 F 0.0  
20:00 192.6 F 0.0  
25:00 190.5 F 0.0  
30:00 187.8 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 22:16  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

07/12/2023 12:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_