
Midmark M11 - v1.0.0
Total Cycles: 3848

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2023 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 201.8 F 3.3
4:00 212.7 F 8.4
6:00 224.8 F 8.3
8:00 232.0 F 9.0
10:00 241.0 F 10.3
12:00 243.6 F 12.4
14:00 242.6 F 11.3
16:00 256.1 F 18.3
18:00 268.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 271.7 F 27.9
2:00 272.0 F 28.1
3:00 272.5 F 28.5
4:00 272.3 F 28.7

Min 271.5 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 217.2 F 0.0
10:00 230.3 F 0.0
15:00 227.8 F 0.0
20:00 225.0 F 0.0
25:00 222.0 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

10/02/2023 12:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____