

-----  
Midmark M11 - v1.0.5  
Total Cycles: 672

Name: RITTER2  
Sterilizer ID: V2313228  
01/05/2022 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.6 F 0.0  
2:00 82.6 F 0.2  
4:00 94.3 F 1.4  
6:00 131.9 F 3.8  
8:00 158.9 F 7.2  
10:00 183.2 F 11.5  
12:00 207.8 F 3.8  
14:00 218.2 F 7.6  
16:00 231.0 F 11.8  
18:00 231.0 F 8.0  
20:00 243.3 F 13.5  
22:00 242.3 F 11.8  
24:00 254.6 F 18.4  
26:00 266.0 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.0 F 29.5  
2:00 272.8 F 29.7  
3:00 272.5 F 29.3  
4:00 272.2 F 29.1

Min 271.6 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 224.1 F 0.0  
5:00 224.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:15  
HEATING: 27:30  
STERILIZING: 4:00  
VENTING: 2:40  
DRYING: 9:13  
TOTAL CYCLE: 00:44:37

01/05/2022 12:44 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_