
Midmark M11 - v1.0.0
Total Cycles: 1120

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2021 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 179.1 F 2.9
4:00 193.5 F 6.4
6:00 207.5 F 10.8
8:00 219.4 F 4.6
10:00 233.9 F 9.6
12:00 236.6 F 9.7
14:00 241.2 F 11.5
16:00 238.7 F 9.6
18:00 251.5 F 15.9
20:00 262.5 F 22.3
22:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 28.9
2:00 271.7 F 28.3
3:00 271.9 F 28.9
4:00 272.4 F 28.8

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 227.5 F 0.0
10:00 224.3 F 0.0
15:00 221.0 F 0.0
20:00 223.2 F 0.0
25:00 225.7 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

08/25/2021 02:14 PM
ACCEPTED: X
REJECTED:
Operator: U02