

-----  
Midmark M11 - v1.0.0  
Total Cycles: 600

Name: MIDMARK  
Sterilizer ID: V2178198  
01/13/2021 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.0 F 0.0  
2:00 167.6 F 2.0  
4:00 191.7 F 6.6  
6:00 213.7 F 12.6  
8:00 224.3 F 7.1  
10:00 238.5 F 13.0  
12:00 239.0 F 10.5  
14:00 236.6 F 8.8  
16:00 251.8 F 16.0  
18:00 264.1 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.0 F 28.3  
2:00 272.7 F 28.7  
3:00 271.8 F 28.4  
4:00 272.1 F 28.3  
  
Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 208.1 F 0.0  
10:00 193.9 F 0.0  
15:00 191.2 F 0.0  
20:00 204.4 F 0.0  
25:00 202.5 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

01/13/2021 09:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01