
Midmark M11 - v1.0.0
Total Cycles: 5364

Name: RITTER 1
Sterilizer ID: V2178198
02/25/2025 02:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.3 F 0.0
2:00 130.6 F 1.4
4:00 138.6 F 4.1
6:00 167.7 F 7.2
8:00 190.8 F 10.9
10:00 208.6 F 4.7
12:00 216.0 F 7.3
14:00 227.5 F 11.6
16:00 227.6 F 7.4
18:00 239.8 F 12.7
20:00 238.5 F 10.6
22:00 250.7 F 16.6
24:00 261.5 F 22.8
26:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 30.0
2:00 272.7 F 29.7
3:00 271.6 F 29.1
4:00 271.7 F 29.6

Min 271.4 F 29.1
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 195.6 F 0.0
10:00 198.6 F 0.0
15:00 204.8 F 0.0
20:00 207.3 F 0.0
25:00 210.8 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:38
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:05:25
CYCLE COMPLETE

02/25/2025 03:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____