

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4294

Name: RITTER 1  
Sterilizer ID: V2178198  
02/12/2024 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.5 F 0.0  
2:00 114.9 F 1.2  
4:00 133.9 F 3.5  
6:00 162.5 F 6.9  
8:00 192.7 F 11.7  
10:00 215.8 F 4.0  
12:00 222.9 F 8.0  
14:00 234.2 F 12.3  
16:00 233.9 F 8.3  
18:00 245.7 F 13.1  
20:00 243.0 F 11.7  
22:00 254.8 F 17.7  
24:00 265.5 F 24.1  
26:00 272.2 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 28.8  
2:00 272.4 F 29.2  
3:00 271.8 F 28.5  
4:00 272.8 F 29.2

Min 271.3 F 28.4  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 232.4 F 0.0  
10:00 223.1 F 0.0  
15:00 230.8 F 0.0  
20:00 211.5 F 0.0  
25:00 214.1 F 0.0  
30:00 200.0 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 26:29  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:04:56  
CYCLE COMPLETE

02/12/2024 03:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_