

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002090  
10:26 21/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:04  
STERILIZING 4:17  
136.4 C 223kPa 4:17  
136.2 C 223kPa 4:47  
136.4 C 223kPa 5:17  
136.2 C 222kPa 5:47  
136.4 C 224kPa 6:17  
136.3 C 225kPa 6:47  
136.4 C 221kPa 7:17  
136.3 C 223kPa 7:47  
136.3 C 223kPa 7:47  
VENTING 7:47  
AIR DRYING 8:39  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
74DD042B6998E110  
-----