
Midmark M11 - v1.0.5
Total Cycles: 2064

Name: RITTER2
Sterilizer ID: V2313228
02/16/2023 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.6 F 0.0
2:00 141.9 F 0.0
4:00 167.3 F 2.5
6:00 191.7 F 6.1
8:00 210.1 F 10.7
10:00 223.6 F 5.5
12:00 236.4 F 11.2
14:00 235.2 F 9.2
16:00 238.4 F 9.7
18:00 249.4 F 15.2
20:00 261.7 F 22.2
22:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.7 F 29.4
2:00 271.5 F 28.7
3:00 272.1 F 29.4
4:00 272.4 F 29.3

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 223.0 F 0.0
10:00 233.2 F 0.0
15:00 230.2 F 0.0
20:00 224.3 F 0.0
25:00 221.2 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

02/16/2023 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____