
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 001323
13:58 27/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:57
STERILIZING 4:01
136.0 C 220kPa 4:01
136.2 C 221kPa 4:31
136.3 C 222kPa 5:01
136.4 C 221kPa 5:31
136.3 C 224kPa 6:01
136.1 C 222kPa 6:31
136.4 C 225kPa 7:01
136.3 C 223kPa 7:31
136.3 C 223kPa 7:31
VENTING 7:31
AIR DRYING 8:25
CYCLE COMPLETE 15:49

Digital Signature #
CBA48F3BAED33E34
