

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3617

Name: RITTER2  
Sterilizer ID: V2313228  
05/14/2024 04:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 172.3 F 3.3  
4:00 194.6 F 8.6  
6:00 214.5 F 7.9  
8:00 233.3 F 9.6  
10:00 235.3 F 8.6  
12:00 244.3 F 12.3  
14:00 245.8 F 13.4  
16:00 259.5 F 20.8  
18:00 271.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 272.5 F 29.0  
2:00 272.9 F 29.6  
3:00 271.8 F 28.6  
4:00 272.9 F 29.3  
  
Min 271.6 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
5:00 223.3 F 0.0  
10:00 211.5 F 0.0  
15:00 211.6 F 0.0  
20:00 210.6 F 0.0  
25:00 207.4 F 0.0  
30:00 202.7 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 18:36  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

05/14/2024 04:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_