
Midmark M11 - v1.0.5
Total Cycles: 1753

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.6 F 0.0
2:00 197.3 F 1.6
4:00 205.0 F 7.3
6:00 225.3 F 13.2
8:00 233.6 F 10.3
10:00 235.3 F 8.7
12:00 239.1 F 9.7
14:00 250.4 F 15.8
16:00 263.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.2 F 29.3
2:00 271.4 F 29.2
3:00 272.2 F 29.4
4:00 271.2 F 28.9

Min 271.2 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 219.1 F 0.0
10:00 233.2 F 0.0
15:00 232.0 F 0.0
20:00 227.4 F 0.0
25:00 221.4 F 0.0
Dry Cycle Aborted

FILLING: 1:46
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 25:16
TOTAL CYCLE: 00:52:13

11/21/2022 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____