

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 000718  
8:37 29/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:19  
STERILIZING 4:16  
136.0 C 220kPa 4:16  
136.4 C 224kPa 4:46  
136.2 C 223kPa 5:16  
136.3 C 222kPa 5:46  
136.4 C 223kPa 6:16  
136.3 C 221kPa 6:46  
136.2 C 223kPa 7:16  
135.8 C 220kPa 7:46  
135.8 C 220kPa 7:46  
VENTING 7:46  
AIR DRYING 8:36  
CYCLE COMPLETE 8:40

-----  
Digital Signature #  
A463C5E373116EFF  
-----