
Midmark M11 - v1.0.0
Total Cycles: 1550

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 05:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 173.7 F 2.9
4:00 178.6 F 7.1
6:00 205.7 F 11.9
8:00 220.7 F 5.5
10:00 232.2 F 10.5
12:00 235.2 F 7.2
14:00 243.8 F 12.9
16:00 242.6 F 11.4
18:00 255.7 F 18.2
20:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.0
3:00 272.7 F 28.9
4:00 272.1 F 28.4

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 223.1 F 0.0
10:00 233.2 F 0.0
15:00 229.2 F 0.0
20:00 228.7 F 0.0
25:00 230.1 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

01/20/2022 06:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____