

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1087

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2021 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.8 F 0.0  
2:00 125.5 F 0.0  
4:00 138.0 F 0.2  
6:00 165.1 F 2.0  
8:00 187.4 F 4.9  
10:00 203.7 F 8.5  
12:00 217.1 F 12.6  
14:00 223.6 F 6.7  
16:00 235.4 F 11.2  
18:00 235.2 F 7.9  
20:00 244.3 F 12.2  
22:00 244.5 F 12.5  
24:00 256.3 F 18.7  
26:00 267.6 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.7 F 28.9  
2:00 271.6 F 28.2  
3:00 272.7 F 29.1  
4:00 272.2 F 28.6

Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 206.2 F 0.0  
10:00 213.2 F 0.0  
15:00 207.4 F 0.0  
20:00 201.0 F 0.0  
25:00 197.3 F 0.0  
30:00 192.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 27:22  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:05:37  
CYCLE COMPLETE

08/10/2021 02:07 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01