

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1489

Name: RITTER2  
Sterilizer ID: V2313228  
09/14/2022 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
2:00 192.2 F 3.2  
4:00 192.6 F 7.9  
6:00 218.4 F 13.5  
8:00 227.2 F 9.3  
10:00 233.7 F 9.0  
12:00 242.6 F 13.3  
14:00 243.4 F 12.7  
16:00 256.4 F 19.6  
18:00 267.9 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.7 F 30.2  
2:00 272.5 F 29.6  
3:00 271.5 F 29.2  
4:00 272.0 F 29.9  
  
Min 271.3 F 29.1  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 221.1 F 0.0  
10:00 241.0 F 0.0  
15:00 231.9 F 0.0  
20:00 227.5 F 0.0  
25:00 225.6 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:25  
CYCLE COMPLETE

09/14/2022 11:11 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_