

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1197

Name: RITTER 1  
Sterilizer ID: V2178198  
09/23/2021 06:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.7 F 0.0  
2:00 78.6 F 0.0  
4:00 92.5 F 0.0  
6:00 115.0 F 0.1  
8:00 144.5 F 0.1  
10:00 173.8 F 0.2  
12:00 195.2 F 1.4  
14:00 210.5 F 4.5  
16:00 228.9 F 7.6  
18:00 236.4 F 11.0  
20:00 231.3 F 7.5  
22:00 240.1 F 11.1  
24:00 234.4 F 8.0  
26:00 243.5 F 11.6  
28:00 246.2 F 13.1  
30:00 258.0 F 19.4  
32:00 268.6 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.6 F 28.8  
2:00 271.7 F 28.0  
3:00 272.0 F 28.7  
4:00 272.2 F 28.5  
  
Min 271.2 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
5:00 225.5 F 0.0  
10:00 222.5 F 0.0  
15:00 224.4 F 0.0  
20:00 218.2 F 0.0  
25:00 231.7 F 0.0  
30:00 220.4 F 0.0  
  
Drying Complete

FILLING: 1:07  
HEATING: 33:13  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:11:36  
CYCLE COMPLETE

09/23/2021 08:10 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_