

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2923

Name: RITTER 1  
Sterilizer ID: V2178198  
02/21/2023 11:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 217.5 F 0.0  
2:00 214.7 F 5.0  
4:00 210.1 F 10.7  
6:00 224.8 F 6.6  
8:00 239.1 F 12.7  
10:00 238.1 F 10.7  
12:00 237.6 F 9.5  
14:00 252.2 F 16.9  
16:00 264.9 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.7  
1:00 272.2 F 29.4  
2:00 272.2 F 29.0  
3:00 271.6 F 28.9  
4:00 272.6 F 29.5  
  
Min 271.4 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 230.1 F 0.0  
10:00 236.4 F 0.0  
15:00 232.2 F 0.0  
20:00 225.0 F 0.0  
25:00 224.7 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 17:39  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:57:06  
CYCLE COMPLETE

02/21/2023 12:46 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_