

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1400

Name: RITTER2  
Sterilizer ID: V2313228  
08/16/2022 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.9 F 0.0  
2:00 172.2 F 0.5  
4:00 198.3 F 5.7  
6:00 215.9 F 11.2  
8:00 226.1 F 7.3  
10:00 239.8 F 12.9  
12:00 240.3 F 11.4  
14:00 240.1 F 10.6  
16:00 254.0 F 17.7  
18:00 265.7 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.2 F 29.1  
2:00 272.7 F 29.6  
3:00 271.7 F 28.7  
4:00 272.7 F 29.4  
  
Min 271.6 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 224.9 F 0.0  
10:00 226.5 F 0.0  
15:00 225.3 F 0.0  
20:00 231.1 F 0.0  
25:00 220.9 F 0.0  
30:00 220.4 F 0.0  
Drying Complete

FILLING: 2:31  
HEATING: 19:38  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

08/16/2022 02:14 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_