

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3451

Name: RITTER2  
Sterilizer ID: V2313228  
03/12/2024 10:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.9 F 0.0  
2:00 128.8 F 0.2  
4:00 136.4 F 1.9  
6:00 167.9 F 5.1  
8:00 187.7 F 9.1  
10:00 205.4 F 13.4  
12:00 218.1 F 7.2  
14:00 230.1 F 11.8  
16:00 232.7 F 8.8  
18:00 239.6 F 11.0  
20:00 245.2 F 13.4  
22:00 256.9 F 19.6  
24:00 267.7 F 26.2  
26:00 271.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 271.5 F 29.3  
2:00 272.3 F 29.6  
3:00 272.6 F 29.9  
4:00 271.4 F 29.9

Min 271.0 F 29.3  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 169.9 F 0.0  
5:00 201.3 F 0.0  
10:00 196.5 F 0.0  
15:00 196.3 F 0.0  
20:00 195.4 F 0.0  
25:00 193.8 F 0.0  
30:00 191.2 F 0.0  
Drying Complete

FILLING: 0:39  
HEATING: 26:05  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:04:21  
CYCLE COMPLETE

03/12/2024 11:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_