

-----  
Midmark M11 - v1.0.5  
Total Cycles: 752

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2022 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.3 F 0.0  
2:00 163.6 F 3.2  
4:00 185.6 F 7.4  
6:00 208.2 F 12.9  
8:00 222.5 F 7.7  
10:00 236.6 F 13.4  
12:00 239.3 F 11.3  
14:00 239.2 F 10.3  
16:00 253.2 F 17.4  
18:00 265.3 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.8 F 29.7  
2:00 272.0 F 29.0  
3:00 271.6 F 29.2  
4:00 272.8 F 29.7  
  
Min 271.4 F 28.7  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
5:00 215.2 F 0.0  
10:00 203.6 F 0.0  
15:00 202.6 F 0.0  
20:00 199.7 F 0.0  
25:00 197.4 F 0.0  
30:00 195.2 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

02/02/2022 02:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_