

-----  
Midmark M11 - v1.0.5  
Total Cycles: 945

Name: RITTER2  
Sterilizer ID: V2313228  
04/01/2022 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 186.6 F 3.1  
4:00 208.2 F 8.9  
6:00 225.2 F 8.7  
8:00 237.0 F 12.1  
10:00 238.7 F 11.0  
12:00 239.1 F 10.6  
14:00 253.8 F 18.0  
16:00 266.3 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.2 F 29.5  
2:00 272.4 F 30.0  
3:00 271.6 F 29.1  
4:00 272.5 F 29.8  
  
Min 271.4 F 29.1  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 231.3 F 0.0  
10:00 231.7 F 0.0  
15:00 225.5 F 0.0  
20:00 224.6 F 0.0  
25:00 227.1 F 0.0  
30:00 230.4 F 0.0  
  
Drying Complete

FILLING: 3:07  
HEATING: 17:29  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:47  
CYCLE COMPLETE

04/01/2022 12:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_