
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005210
9:42 13/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:23
136.0 C 220kPa 3:23
136.1 C 221kPa 3:53
136.2 C 222kPa 4:23
136.5 C 224kPa 4:53
136.4 C 223kPa 5:23
136.2 C 223kPa 5:53
136.4 C 221kPa 6:23
136.4 C 223kPa 6:53
136.4 C 223kPa 6:53
VENTING 6:53
AIR DRYING 7:42
CYCLE COMPLETE 10:30

Digital Signature #
0B613DA6039EF0F6
