

-----  
Midmark M11 - v1.0.5  
Total Cycles: 275

Name: RITTER2  
Sterilizer ID: V2313228  
08/10/2021 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.0 F 0.0  
2:00 120.1 F 0.7  
4:00 135.6 F 3.1  
6:00 150.8 F 6.9  
8:00 177.5 F 11.5  
10:00 221.3 F 4.0  
12:00 226.6 F 8.9  
14:00 231.3 F 11.9  
16:00 240.3 F 10.8  
18:00 235.8 F 8.8  
20:00 250.0 F 15.3  
22:00 261.1 F 21.7  
24:00 271.6 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 272.2 F 29.3  
2:00 272.0 F 28.9  
3:00 271.6 F 29.3  
4:00 272.1 F 29.4

Min 271.3 F 28.9  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
5:00 233.8 F 0.0  
10:00 232.5 F 0.0  
15:00 226.1 F 0.0  
20:00 225.0 F 0.0  
Dry Cycle Aborted

FILLING: 1:00  
HEATING: 25:09  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 24:53  
TOTAL CYCLE: 00:57:57

08/10/2021 09:33 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01