

-----  
Midmark M11 - v1.0.5  
Total Cycles: 718

Name: RITTER2  
Sterilizer ID: V2313228  
01/20/2022 04:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 124.0 F 0.0  
2:00 129.4 F 1.8  
4:00 153.3 F 5.1  
6:00 181.1 F 9.7  
8:00 207.2 F 6.3  
10:00 220.8 F 8.4  
12:00 233.5 F 12.2  
14:00 237.4 F 10.9  
16:00 236.9 F 9.4  
18:00 250.9 F 16.2  
20:00 262.8 F 23.2  
22:00 272.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.7 F 29.8  
2:00 271.8 F 29.1  
3:00 271.8 F 29.5  
4:00 272.9 F 29.8

Min 271.4 F 28.9  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 201.2 F 0.0  
10:00 206.1 F 0.0  
15:00 210.2 F 0.0  
20:00 212.7 F 0.0  
25:00 216.2 F 0.0  
30:00 217.7 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 22:14  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:00:47  
CYCLE COMPLETE

01/20/2022 05:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_